



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Conductivity Calibration Report

Customer:	EMS/VOS Nippon		
Job Number:	74141	Date of Report:	6/6/2013
Model Number	SBE 45	Serial Number:	4525798-0028

*Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.*

*An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.*

### 'AS RECEIVED CALIBRATION'

☐ Performed ☒ Not Performed

Date:

Drift since last cal:  PSU/month

Comments:

### 'CALIBRATION AFTER REPAIR'

☒ Performed ☐ Not Performed

Date:  6/6/2013

Drift since Last cal:  N/A PSU/month

Comments:

The board set and conductivity cell were replaced.

*\*Measured at 3.0 S/m*

*Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.*

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0028  
CALIBRATION DATE: 06-Jun-13

SBE 45 CONDUCTIVITY CALIBRATION DATA  
PSS 1978:  $C(35,15,0) = 4.2914$  Siemens/meter

## COEFFICIENTS:

$g = -9.863054e-001$   
 $h = 1.294409e-001$   
 $i = -3.305103e-004$   
 $j = 4.212423e-005$

$CPcor = -9.5700e-008$   
 $CTcor = 3.2500e-006$   
 $WBOTC = 1.0240e-006$


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2766.69	0.00000	0.00000
0.9999	34.8038	2.97499	5543.54	2.97500	0.00001
4.5000	34.7842	3.28199	5753.67	3.28199	-0.00001
15.0000	34.7418	4.26344	6378.24	4.26344	-0.00000
18.5000	34.7328	4.60849	6583.41	4.60848	-0.00000
24.0000	34.7228	5.16625	6901.80	5.16627	0.00001
29.0000	34.7173	5.68793	7186.39	5.68792	-0.00001

$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$







# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Conductivity Calibration Report

Customer:	EMS/VOS Nippon		
Job Number:	68841	Date of Report:	5/23/2012
Model Number	SBE 45	Serial Number:	4525798-0028

Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.

### 'AS RECEIVED CALIBRATION'

☐ Performed ☒ Not Performed

Date:

Drift since last cal:  PSU/month\*

Comments:

### 'CALIBRATION AFTER CLEANING & REPLATINIZING'

☒ Performed ☐ Not Performed

Date:

Drift since Last cal:  PSU/month\*

Comments:

*\*Measured at 3.0 S/m*

*Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.*

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0028  
CALIBRATION DATE: 22-May-12

SBE 45 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## COEFFICIENTS:

g = -1.034520e+000  
h = 1.445394e-001  
i = -2.976732e-004  
j = 4.409474e-005

CPcor = -9.5700e-008  
CTcor = 3.2500e-006  
WBOTC = 6.3756e-006

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2679.60	0.00000	0.00000
1.0000	34.9130	2.98344	5278.69	2.98347	0.00003
4.5000	34.8927	3.29122	5476.38	3.29118	-0.00004
15.0000	34.8496	4.27527	6064.70	4.27528	0.00001
18.5000	34.8401	4.62118	6258.06	4.62119	0.00001
24.0000	34.8297	5.18040	6558.27	5.18040	0.00000
29.0000	34.8228	5.70326	6826.65	5.70324	-0.00002
32.5000	34.8172	6.07613	7011.56	6.07614	0.00001


$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

Conductivity =  $(g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p)$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$ ;

Residual = instrument conductivity - bath conductivity

Date, Slope Correction



● 30-Jul-11 1.0016784  
▲ 22-May-12 1.0000000

**CALIBRATION AFTER  
CLEANING AND  
REPLATINIZING CELL**



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Conductivity Calibration Report

Customer:	EMS/VOS Nippon		
Job Number:	65183	Date of Report:	8/1/2011
Model Number	SBE 45	Serial Number:	4525798-0028

*Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.*

*An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients using the program SEACON. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.*

### 'AS RECEIVED CALIBRATION'

☒ Performed ☐ Not Performed

Date: 7/30/2011

Drift since last cal: +0.00070 PSU/month\*

Comments:

### 'CALIBRATION AFTER CLEANING & REPLATINIZING'

☐ Performed ☒ Not Performed

Date:

Drift since Last cal: PSU/month\*

Comments:

*\*Measured at 3.0 S/m*

*Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.*

# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0028  
CALIBRATION DATE: 30-Jul-11

SBE 45 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## COEFFICIENTS:

g = -1.034877e+000  
h = 1.442621e-001  
i = -1.431364e-004  
j = 3.189645e-005

CPcor = -9.5700e-008  
CTcor = 3.2500e-006  
WBOTC = 6.3756e-006

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2679.60	0.00000	0.00000
1.0000	34.8692	2.98006	5273.06	2.98005	-0.00001
4.5000	34.8492	3.28752	5470.48	3.28753	0.00001
15.0000	34.8065	4.27054	6057.80	4.27054	-0.00000
18.5000	34.7976	4.61616	6250.93	4.61616	0.00000
24.0000	34.7875	5.17482	6550.76	5.17482	0.00000
29.0000	34.7810	5.69719	6818.87	5.69717	-0.00002
32.5000	34.7743	6.06949	7003.53	6.06951	0.00001


$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

Date, Slope Correction



**POST CRUISE  
CALIBRATION**





# SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Conductivity Calibration Report

Customer:	EMS/VOS Nippon		
Job Number:	55277	Date of Report:	7/29/2009
Model Number:	SBE 45	Serial Number:	4525798-0028

Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients using the program SEACON. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.

### 'AS RECEIVED CALIBRATION'

☒ Performed ☐ Not Performed

Date: 7/17/2009

Drift since last cal: +0.00050 PSU/month\*

Comments:

### 'CALIBRATION AFTER CLEANING & REPLATINIZING'

☒ Performed ☐ Not Performed

Date: 7/28/2009

Drift since 08 Sep 07 -0.00010 PSU/month\*

Comments:

*\*Measured at 3.0 S/m*

*Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.*



# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0028  
CALIBRATION DATE: 17-Jul-09

SBE 45 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## COEFFICIENTS:

g = -1.034472e+000  
h = 1.442046e-001  
i = -1.368575e-004  
j = 3.252687e-005

CPcor = -9.5700e-008  
CTcor = 3.2500e-006  
WBOTC = 6.3756e-006

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREO (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2679.42	0.00000	0.00000
0.9999	34.9432	2.98577	5276.66	2.98576	-0.00001
4.5000	34.9218	3.29369	5474.19	3.29370	0.00001
15.0000	34.8782	4.27840	6061.96	4.27840	0.00000
18.5000	34.8690	4.62460	6255.21	4.62461	0.00000
24.0000	34.8587	5.18423	6555.24	5.18423	-0.00000
29.0000	34.8527	5.70761	6823.57	5.70760	-0.00001
32.5000	34.8487	6.08100	7008.54	6.08101	0.00001


$$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

Date, Slope Correction



**POST CRUISE  
CALIBRATION**





# SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Conductivity Calibration Report

Customer:	EMS/VOS Nippon		
Job Number:	48029	Date of Report:	9/10/2007
Model Number	SBE 45	Serial Number:	4525798-0028

Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients using the program SEACON. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.

### 'AS RECEIVED CALIBRATION'

☒ Performed ☐ Not Performed

Date: 9/8/2007

Drift since last cal: +0.00070 PSU/month\*

Comments:

### 'CALIBRATION AFTER CLEANING & REPLATINIZING'

☐ Performed ☒ Not Performed

Date:

Drift since Last cal: PSU/month\*

Comments:

*\*Measured at 3.0 S/m*

*Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.*



# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0028  
CALIBRATION DATE: 08-Sep-07

SBE 45 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## COEFFICIENTS:

g = -1.035560e+000  
h = 1.444473e-001  
i = -1.891727e-004  
j = 3.639980e-005

CPcor = -9.5700e-008  
CTcor = 3.2500e-006  
WBOTC = 6.3756e-006

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2679.61	0.00000	0.00000
1.0000	34.7879	2.97377	5268.18	2.97378	0.00001
4.4999	34.7679	3.28060	5465.24	3.28058	-0.00001
15.0000	34.7248	4.26157	6051.61	4.26157	-0.00000
18.5000	34.7157	4.60646	6244.39	4.60646	0.00000
24.0000	34.7055	5.16396	6543.71	5.16397	0.00001
29.0000	34.6994	5.68532	6811.37	5.68532	-0.00000
32.5000	34.6951	6.05724	6995.86	6.05724	-0.00000


$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

Date, Slope Correction



● 09-Aug-05 1.0005913  
▲ 08-Sep-07 1.0000000

**POST CRUISE  
CALIBRATION**





## SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

### Conductivity Calibration Report

Customer:	EMS/VOS Nippon		
Job Number:	40269	Date of Report:	8/9/2005
Model Number:	SBE 45	Serial Number:	4525798-0028

Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients using the program SEACON. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.

#### 'AS RECEIVED CALIBRATION'

☒ Performed ☐ Not Performed

Date: 8/3/2005

Drift since last cal: +.00250 PSU/month\*

#### Comments:

The conductivity cell was found to have failed.

#### 'CALIBRATION AFTER REPAIR'

☒ Performed ☐ Not Performed

Date: 8/9/2005

Drift since Last cal: N/A PSU/month\*

#### Comments:

The conductivity cell was replaced.

*\*Measured at 3.0 S/m*

*Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.*

**SEA-BIRD ELECTRONICS, INC.**  
 1808 136th Place N.E., Bellevue, Washington, 98005 USA  
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0028  
 CALIBRATION DATE: 02-Aug-05

SBE 45 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.722572e-001  
 h = 1.359101e-001  
 i = -4.853963e-004  
 j = 5.302932e-005

CPcor = -9.5700e-008  
 CTcor = 3.2500e-006  
 WBOTC = 6.3756e-006


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2683.55	0.00000	0.00000
1.0000	34.8116	2.97560	5410.98	2.97560	-0.00000
4.4999	34.7917	3.28263	5616.99	3.28263	0.00000
15.0000	34.7488	4.26421	6229.17	4.26422	0.00002
18.4999	34.7400	4.60933	6430.22	4.60932	-0.00002
24.0000	34.7302	5.16723	6742.22	5.16724	0.00001
29.0000	34.7258	5.68916	7021.12	5.68916	-0.00001
32.5000	34.7241	6.06173	7213.37	6.06174	0.00000

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$



**POST CRUISE  
 CALIBRATION**





## SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

### Conductivity Calibration Report

Customer:	EMS/VOS Nippon		
Job Number:	33506	Date of Report:	9/23/2003
Model Number:	SBE 45	Serial Number:	4525798-0028

Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients using the program SEACON. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.

#### 'AS RECEIVED CALIBRATION'

☒ Performed ☐ Not Performed

Date: 9/12/2003

Drift since last cal: +.00140 PSU/month\*

#### Comments:

The conductivity cell was found to have failed.

#### 'CALIBRATION AFTER CLEANING & REPLATINIZING'

☒ Performed ☐ Not Performed

Date: 9/23/2003

Drift since Last cal: N/A PSU/month\*

#### Comments:

The conductivity cell was replaced.

*\*Measured at 3.0 S/m*

*Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.*

**SEA-BIRD ELECTRONICS, INC.**  
 1808 136th Place N.E., Bellevue, Washington, 98005 USA  
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0028  
 CALIBRATION DATE: 02-Aug-05

SBE 45 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.722572e-001  
 h = 1.359101e-001  
 i = -4.853963e-004  
 j = 5.302932e-005

CPcor = -9.5700e-008  
 CTcor = 3.2500e-006  
 WBOTC = 6.3756e-006


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2683.55	0.00000	0.00000
1.0000	34.8116	2.97560	5410.98	2.97560	-0.00000
4.4999	34.7917	3.28263	5616.99	3.28263	0.00000
15.0000	34.7488	4.26421	6229.17	4.26422	0.00002
18.4999	34.7400	4.60933	6430.22	4.60932	-0.00002
24.0000	34.7302	5.16723	6742.22	5.16724	0.00001
29.0000	34.7258	5.68916	7021.12	5.68916	-0.00001
32.5000	34.7241	6.06173	7213.37	6.06174	0.00000

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$



**POST CRUISE  
 CALIBRATION**